

# Ultrasonic Contact Impedance (UCI) accuracy, measurement deviation and repeatability explained

Many producers talk about the UCI accuracy and how accurate the probes are, but this creates even greater confusion among the users when it comes to the technique itself.

**So why is the probe accuracy an inaccurate statement?** The probe accuracy refers to “how accurate the technology and its components” are but does not define what multi-point accuracy the probe can deliver. Even more important is that the UCI method is strongly influenced by other parameters, such as operator’s experience or handling. For the UCI method the two other parameters are of much greater significance for the user, when it comes to device performance: **measurement deviation** and **repeatability**.

**So how these two parameters are described and what do they really mean?**

According to DIN 50159 and A1038 the **measurement deviation** is defined as follows:

$$E = \frac{\bar{H} - H}{H} \cdot 100 \%$$

(E – measurement deviation,  $\bar{H}$  – Average value of n measurements, H – reference value i.e. test block)

In other words: it describes how the average value deviates from the reference value.

**The repeatability of UCI** also denoted as coefficient of variation is defined in **DIN 50159** and describes the relative difference between the highest and the lowest hardness value with respect to the average:

$$r = \frac{H_{max} - H_{min}}{\bar{H}} \cdot 100 \%$$

(r-repeatability,  $H_{min}$  and  $H_{max}$  – the lowest and the highest hardness values respectively,  $\bar{H}$ - average value)

In other words: it describes how far the measurement values are scattered from each other.

**So what are the numbers I should pay attention to?**

Let the following Table 1 be a quick guideline.

Scale	Max. measurement deviation (E) in % DIN 50159 & ASTM A1038								Repeatability %	
	<250 HV		250 -500 HV		500 – 800 HV		>800 HV		≤ 250 HV	> 250 HV
	DIN	ASTM	DIN	ASTM	DIN	ASTM	DIN	ASTM	DIN	DIN
HV 0.1	5	6	6	7	7	8	8	9	8	6
HV 0.3	5	6	6	7	7	8	8	9	8	6
HV 0.8	4	6	4	7	5	8	6	9	8	6
HV 1	4	5	4	5	5	7	6	7	8	6
HV 5	4	5	4	5	4	7	4	7	5	5
HV 10	4	5	4	5	4	7	4	7	5	5

Table 1. The summary of maximum tolerable errors for measurement deviation and repeatability from DIN 50159 and ASTM A1038. The values are adopted from DIN 50159 and ASTM A1038.

**Why there are different values for different scales and hardness ranges ?**

The UCI method uses a vibrating rod (with Vickers diamond at the tip) to measure the indentation depth. The indenter vibrates with a specific frequency ( $f_e$ ), which changes upon the indentation ( $f_i$ ) or better to say as the response to the contact area between the diamond and the material. The greater the indentation the higher the frequency change. Typically, the indentation depth varies between 5  $\mu\text{m}$  and 35  $\mu\text{m}$ .

It is correct to say that the greater frequency changes can be measured more accurately, indicating: the higher the contact with the diamond (i.e. higher indentation depth) the lower the uncertainty of the measurement.

Higher tolerance values for lower HV scales and harder material ranges are based on the fact that for the low HV scales such as HV 0.1- HV 0.8 the low force is applied, leading to lower material penetration by the intender.

This combined with other measurement uncertainties such as different than 90° angle of the probe or wobbly hand of the operator can contribute additionally to a greater measurement deviation. This effect is even more pronounced for harder materials, whereby the penetration depth is even lower. In other words: the highest accuracy is expected for soft materials measured with HV 10 scale whereas hard materials measured with the low force are more prone to errors.

**Does the calibration certificate include all necessary values ?**

Yes, UCI probes from Proceq comply to both DIN 50159 and ASTM A1038. Measurement deviation and repeatability are specified on each certificate delivered with each UCI probe.

<b>Object</b>	<b>Equotip UCI Probe ET50-006-0214</b>	<b>equotip<sup>®</sup></b>																									
<b>On basis of</b>	DIN 50159 and ASTM A1038																										
<b>Type of verification</b>	Indirect - Single level																										
<b>Verification equipment</b>	UP01-003-1157																										
<b>Reference test block</b>	HV85-015-0010 HV1 HV85-015-0010 HV5 HV85-101-0014 HV10	<b>Hardness values of 5 readings</b> <table border="1"> <thead> <tr> <th>No</th> <th>1</th> <th>2</th> <th>3</th> <th>4</th> <th>5</th> </tr> </thead> <tbody> <tr> <td>HV1(UCI)</td> <td>856.0</td> <td>853.0</td> <td>843.0</td> <td>853.0</td> <td>856.0</td> </tr> <tr> <td>HV5(UCI)</td> <td>849.0</td> <td>849.0</td> <td>844.0</td> <td>846.0</td> <td>850.0</td> </tr> <tr> <td>HV10(UCI)</td> <td>836.0</td> <td>847.0</td> <td>837.0</td> <td>842.0</td> <td>843.0</td> </tr> </tbody> </table>		No	1	2	3	4	5	HV1(UCI)	856.0	853.0	843.0	853.0	856.0	HV5(UCI)	849.0	849.0	844.0	846.0	850.0	HV10(UCI)	836.0	847.0	837.0	842.0	843.0
No	1	2	3	4	5																						
HV1(UCI)	856.0	853.0	843.0	853.0	856.0																						
HV5(UCI)	849.0	849.0	844.0	846.0	850.0																						
HV10(UCI)	836.0	847.0	837.0	842.0	843.0																						
<b>Reference value</b>	848.0 HV1 846.0 HV5 843.0 HV10	<b>Coefficient of variation (repeatability)</b> HV1: 1.5 % HV5: 0.7 % HV10: 1.3 %																									
<b>Enlarged uncertainty of</b>		<b>Measurement deviation</b> HV1: 0.5 % HV5: 0.2 % HV10: 0.2 %																									
- Verification machine $U_{CM}$	1.0 HV	<b>Measurement result on reference</b> HV1: 852.2 HV5: 847.6 HV10: 841.0																									
- Hardness test block $U_{RM}(k=2)$	17.7 HV1 8.8 HV5 9.6 HV10																										
<b>Temperature</b>	23.0 $\pm$ 2°C																										
<b>Date of verification</b>	2019-08-19	<b>Verification result</b> Within tolerance																									
Verification shall be within 12 months as specified in DIN 50159-2.		<b>Inspector</b>	P. Zioerjen _____																								
<b>Next verification due</b>	2020-08-18	The device specified was subjected to inspection by a trained personnel using software and equipment designed and supplied by Proceq SA. This instrument meets all product specifications and quality requirements with respect to measurement criteria. Proceq SA quality management system is certified according to ISO 9001																									
The device performance is verified using test procedures in compliance with DIN 50159-2 and ASTM A1038 standards. Tools used for the verification are calibrated and traceable to PTB, NIST and/or manufacturer's reference standards.																											

Fig.1. Example of UCI certificate issued by Proceq.

**Disclaimer:**

This document shows only a fraction of the information described in DIN 50159 and ASTM A1038. Proceq AG has done everything in its power to translate accurately the sections of the DIN 50159 Standard. For an authorized translation of the DIN 50159 standard or more information the interested readers are encouraged to read the full version of standards DIN 50159 and ASTM A1038 available at [www.beuth.de](http://www.beuth.de) and [www.astm.org](http://www.astm.org).