



Test Products International, Inc.

9615 SW Allen Blvd. Suite 104
Beaverton, OR 97005

Ph. 503-520-9197
Fax. 503-520-1225

Certificate of Compliance

Certificate Number: CBM: 3513

Date:: 3/20/2024

Instrument Information

Model Number: 9071

Serial Number: 26946750047

Description: Vibration meter

Certificate Information

Adjustments Made: Full Calibration

Calibration Date: 3/20/2024

***Calibration Due Date:** 3/20/2025
or 12 months
from first use.

**Basic
Accuracy:** 5%

Procedure: 9071 CAL

Cal By: JC

Test Products International, Inc. certifies the above instrumentation has been calibrated to meet or exceed published specifications in accordance with the above referenced procedure.

* Many factors may cause this instrument to drift out of calibration prior to the recommended calibration due date.

Calibration Standard Used

Manufacturer: MMF

Model: VC20

Description: Vibration Calibrator

Serial Number: 130400

Cert ID: 50528

Due Date: 11/9/24

Technical Manager
TPI Technical Department