



# Test Products International, Inc.

9615 SW Allen Blvd. Suite 104  
Beaverton, OR 97005

Ph. 503-520-9197  
Fax. 503-520-1225

## Certificate of Compliance

**Certificate Number:** CBM 1460

**Date::** 7/2/2019

### Instrument Information

<b>Model Number:</b> 9080	<b>Serial Number:</b> 1918T0026	<b>Description:</b> Vibration meter
<b>Model Number:</b> A9012	<b>Serial Number:</b> 492998	<b>Description:</b> Accelerometer

### Certificate Information

<b>Asjustments Made:</b> Full Calibration	<b>Calibration Date:</b> 7/2/2019	<b>*Calibration Due Date:</b> 7/2/2020
<b>Basic Accuracy:</b> 5%	<b>Procedure:</b> 9080 CAL	<b>Cal By:</b> JC

Test Products International, Inc. certifies the above instrumentation has been calibrated to meet or exceed published specifications in accordance with the above referenced procedure.

\* Many factors may cause this instrument to drift out of calibration prior to the recommended calibration due date.

### Calibration Standard Used

<b>Manufacturer:</b> MMF	<b>Model:</b> VC20	<b>Description:</b> Vibration Calibrator
<b>Serial Number:</b> 130400	<b>Cert ID:</b> 40740	

Technical Manager  
TPI Technical Department