



Test Products International, Inc.

9615 SW Allen Blvd. Suite 104
Beaverton, OR 97005

Ph. 503-520-9197
Fax. 503-520-1225

Certificate of Compliance

Certificate Number: CBM 2391

Date: 9/23/2021

Instrument Information

Model Number: 9080	Serial Number: 2133T0030	Description: Vibration meter
Model Number: A9014	Serial Number: 761123	Description: Accelerometer

Certificate Information

Asjustments Made: Full Calibration	Calibration Date: 9/23/2021	*Calibration Due Date: 9/23/2022 or 12 months from first use.
Basic Accuracy: 5%	Procedure: 9080 CAL	Cal By: JC

Test Products International, Inc. certifies the above instrumentation has been calibrated to meet or exceed published specifications in accordance with the above referenced procedure.

* Many factors may cause this instrument to drift out of calibration prior to the recommended calibration due date.

Calibration Standard Used

Manufacturer: MMF	Model: VC20	Description: Vibration Calibrator
Serial Number: 130400	Cert ID: 46576	

Technical Manager
TPI Technical Department