

PERFORMANCE REPORT

PRODUCT MS-100 **DATE OF PRODUCTION** 05/03/2023

STATEMENT OF TESTING The products mentioned on the following page have been thoroughly tested, validated and meet performance accuracy specifications over the stated ranges.

REFERENCE INSTRUMENT Pressure Calibration Type PACE5000
Equipment Accuracy 0.02%RDG+0.02%FS

TEST

PARAMETER

Low pressure accuracy ±8psi			
Standard Value	Display Value	Indication Error	Pass/Fail
50psi	51psi	1psi	pass
100psi	100psi	0psi	pass
200psi	199psi	-1psi	pass
400psi	398psi	-2psi	pass
600psi	598psi	-2psi	pass

High pressure accuracy ±8psi			
Standard Value	Display Value	Indication Error	Pass/Fail
50psi	49psi	-1psi	pass
100psi	100psi	0psi	pass
200psi	199psi	-1psi	pass
400psi	399psi	-1psi	pass
600psi	600psi	0psi	pass

INSPECT BY

Jack Jia