



# Test Products International, Inc.

9615 SW Allen Blvd. Suite 104  
Beaverton, OR 97005

Ph. 503-520-9197  
Fax. 503-520-1225

## Certificate of Compliance

Customer: CULMI AIR-COND & REFRIGERATION PARTS SUP

Certificate ID: 34835

Model Number: 708

Serial Number: 13341040029

Calibration Date: 11/15/2023

\*Calibration Due Date: 11/15/2024

Calibration is due one year from the calibration date. Calibration is recommended annually.

\* Many factors may cause this instrument to drift out of calibration prior to the recommended calibration due date.

Test Products International, Inc. certifies the above instrument has been calibrated to meet or exceed published specifications. This calibration was performed using the standards listed below.

- Calibration Gas: 100ppm Carbon Monoxide / Balance Air
- Calibration Gas: 1000ppm Carbon Monoxide / 3% Oxygen / Balance Nitrogen
- Calibration Gas: 502ppm Nitric Oxide / Balance Nitrogen
- Calibration Gas: 100ppm Nitrogen Dioxide / Balance Nitrogen
- Calibration Gas: 5000ppm Carbon Monoxide Balance Nitrogen / 5% Carbon Monoxide Balance Nitrogen
- Temperature Calibration: 5500A-SC300 Multifunction Calibrator
- Pressure Calibration: 718-30G Pressure Calibrator
- Pressure Calibration: 719-100G Pressure Calibrator
- Calibration Gas: Nitrogen
- Calibration Gas: 5000ppm Methane
- Calibration Gas: 1000ppm Carbon Dioxide / Balance Air
- Temperature and Humidity: Thunder Scientific 1200
- Refrigerant: R134A 0.1 oz/yr Leak Standard
- Vacuum Standard: 25, 250, and 500 microns

To obtain calibration and service of your analyzer, please return it to:

TPI / Attn: Repair  
9615 SW Allen Blvd. Suite 104  
Beaverton, OR 97005

Please include your return / contact information and a description of the service required. For technical assistance call 503-520-9197 or email [tech@testproductsintl.com](mailto:tech@testproductsintl.com)

Test Products International, Inc  
Technical Department