



The Compliance Catalyst



Certificate of Analysis

Lab Ref No.: **PL2008-C27204**

SN: **RS4522466740635061**

Page 1 / 1

Ooi Kah Wai

Chemist

IKM No.: L/2452/7352/16

AN1: RA1501232232206560

DURIO PPE SDN. BHD.

NO 16, JALAN TEMENGGONG,

81100 JOHOR BAHRU JOHOR

MALAYSIA

Date of Issue: 2020-08-19

Date of Received: 2020-08-19

Date of Completion: 2020-08-19

Sample Marking: Durio 501 Adult 3 Ply Disposable Protective Face Mask

Remark: 1) The specification refer to standard specification for performance of materials used in medical face masks ASTM F2100-19.

2)The result recorded pass from 32 out of 32 mask at 160 mmHg.

Test Description	Unit	Result(s)	Spec	Method or Equipment Used
Resistance of Medical Face Mask to Penetration by Synthetic Blood*	mm Hg	160	160	ASTM F1862-07

– END OF REPORT –

This Test Report is prepared according to our standard testing requirements and based solely on the samples submitted by you. We have made every reasonable effort to ensure the accuracy and reliability of the information provided herein. However, we shall not be deemed to have warranted or guaranteed the accuracy and reliability of the results contained in the Test Report under whatsoever circumstances. Neither shall we be liable and responsible for any loss or damage of whatever nature (direct, indirect, incidental, economical, punitive and consequential) caused or alleged to be caused by or in connection with and arising out of the use of or reliance on this Test Report.

