



The Compliance Catalyst



Certificate of Analysis

Lab Ref No.: **PL2007-C24094**

SN: **RS1367902734254016**

Page 1 / 1

Kee Kai Loon
 Microbiologist & Biotechnologist
 Food Analyst No: MJMM 0469
 AN1: RA6338351049449030

DURIO PPE SDN. BHD.
 NO 16, JALAN TEMENGGONG,
 81100 JOHOR BAHRU JOHOR
 MALAYSIA

Date of Issue: 2020-07-28
 Date of Received: 2020-07-22
 Date of Completion: 2020-07-28

Sample Marking: Durio 501 Adult 3 Ply Disposable Protective Face Mask

Challenge Flow Rate: 28.3 L/min
 Challenge level: 4.2×10^5 cfu
 Mean Particle Size: $\sim 3.0 \mu\text{m}$
 Positive control: 2428 cfu

Total CFU Recovered		Bacterial Filtration Efficiency %
Sample 1	ND < 1	> 99.9
Sample 2	ND < 1	
Sample 3	ND < 1	
Sample 4	ND < 1	
Sample 5	ND < 1	

ND denotes Not Detected

Test Description	Unit	Result(s)	Method or Equipment Used
Bacterial Filtration Efficiency*		Refer Above	ASTM F2101

– END OF REPORT –

This Test Report is prepared according to our standard testing requirements and based solely on the samples submitted by you. We have made every reasonable effort to ensure the accuracy and reliability of the information provided herein. However, we shall not be deemed to have warranted or guaranteed the accuracy and reliability of the results contained in the Test Report under whatsoever circumstances. Neither shall we be liable and responsible for any loss or damage of whatever nature (direct, indirect, incidental, economical, punitive and consequential) caused or alleged to be caused by or in connection with and arising out of the use of or reliance on this Test Report.

