



The Compliance Catalyst



## Certificate of Analysis

Lab Ref No.: **PL2008-C26800**

SN: **RS1377126827238806**

Page 1 / 1

**Kee Kai Loon**  
 Microbiologist & Biotechnologist  
 Food Analyst No: MJMM 0469  
 AN1: RA1132195783758474

**DURIO PPE SDN. BHD.**  
 NO 16, JALAN TEMENGGONG,  
 81100 JOHOR BAHRU JOHOR  
 MALAYSIA

Date of Issue: 2020-08-14  
 Date of Received: 2020-08-05  
 Date of Completion: 2020-08-14

Sample Marking: Durio 503 Kid's 3 Ply Disposable Protective Face Mask

Weigh of face mask: 18.31g  
 Volume of extraction: 300 ml  
 TSA count: 54 cfu/100ml  
 SDA count: 1 cfu/100ml

Microbial cleanliness :  $[(54 + 1) (300\text{ml}/100\text{ml})] / 18.31\text{g} = 9 \text{ cfu/g}$

Test Description	Unit	Result(s)	Method or Equipment Used
Microbial cleanliness*	cfu/g	9	ISO 11737-1: 2006

### – END OF REPORT –

This Test Report is prepared according to our standard testing requirements and based solely on the samples submitted by you. We have made every reasonable effort to ensure the accuracy and reliability of the information provided herein. However, we shall not be deemed to have warranted or guaranteed the accuracy and reliability of the results contained in the Test Report under whatsoever circumstances. Neither shall we be liable and responsible for any loss or damage of whatever nature (direct, indirect, incidental, economical, punitive and consequential) caused or alleged to be caused by or in connection with and arising out of the use of or reliance on this Test Report.

