

## PEGREGATE & ROCK Civil Engineering Testing Equipment

www.nl-test.com



## Compression Strength

The hammer is used for rock classification test. Simple and easy to use and is similar to used for concrete test. The level of impact energy is 0.074 kgm percussion energy. Specimen of rock cores is positional horizontally and the rebound index is obtained from the average of several measurements performed perpendicularly to the longitudinal axis.

## **Technical Specifications:**

| Model Number           | NL 1033 X / 001              |
|------------------------|------------------------------|
| Approx. Weight         | 0.75 kg                      |
| Dimension Of Case      | 350 (L) x 85 (W) x 85 (H) mm |
| Level Of Impact Energy | 0.074 kgm                    |

## **Optional Accessories:**

| Model Number           | Accessories Description |
|------------------------|-------------------------|
| NL 1033 X / 001 – A001 | Rock Cradle             |



NL 1033 X / 001 - A001





Copyright 2016 NL Scientific Instruments Sdn. Bhd. All Rights Reserved. Revision 3(All information in this catalogue only for customer's reference and not for quotation. Design and dimension maybe change without prior notice.)