



**Withdraw Tube**



**Filter Funnel**



**Saybolt Thermometer**

## SAYBOLT VISCOMETER APPARATUS

Standard: ASTM D88, AASHTO T72

### Saybolt Universal

Used to determine the saybolt viscosities of petroleum products at the required temperatures between 70°F to 210°F (21.1°C to 98.9°C).

### Technical Specifications :

<b>Model Number</b>	<b>NL 2027 X / 001</b>
Dimension (mm)	270 (L) x 270 (W) x 550 (H)
Approx. Weight	15 kg
Power	240 V, 1 ph, 50 / 60 Hz

### Unit Consists Of :

Model Number	Parts Description	Qty
NL 2027 X / 001 – P 001	Viscometer Probe	2nos.
NL 2027 X / 001 – P 002	Furoal Orifice	2nos.
NL 2027 X / 001 – P 003	Universal Orifice	2nos.
NL 2027 X / 001 – P 004	Flask Pedestal	2nos.
NL 2027 X / 001 – P 005	Temperature Sensor Probe	1 no.
NL 2027 X / 001 – P 006	60 ml Saybolt Viscosity Flask	2nos.

\*1 Copy of Manual Instructions

### Optional Accessories :

NL 2027 X / 001 – A 001	Saybolt Thermometer ASTM 17C, 19°C to 27°C X 0.1°C
NL 2027 X / 001 – A 002	Saybolt Thermometer ASTM 18C, 34°C to 42°C X 0.1°C
NL 2027 X / 001 – A 003	Saybolt Thermometer ASTM 19C, 49°C to 57°C X 0.1°C
NL 2027 X / 001 – A 004	Saybolt Thermometer ASTM 20C, 57°C to 65°C X 0.1°C
NL 2027 X / 001 – A 005	Saybolt Thermometer ASTM 21C, 79°C to 87°C X 0.1°C
NL 2027 X / 001 – A 006	Saybolt Thermometer ASTM 22C, 95°C to 103°C X 0.1°C
NL 2027 X / 001 – A 007	Filter Funnel
NL 2027 X / 001 – A 008	Withdraw Tube
NL 2027 X / 001 – A 009	Bath Oil(5 litres)

\*Manual Instruction 1 copy.