

SAMPLE

検査成績表
<INSPECTION DATA SHEET>

品名<Model Name> (計装ロガー <INSTRUMENTATION LOGGER>)
 形名<Model Number> (LR5031)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Inspection Date> (2011-07-01)
 (<YYYY-MM-DD>)
 検査条件<Inspection Condition> (23.0 °C, 50 %rh)

項目 <Item>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
--------------	---------------	---------------------	--------------------------

電流測定精度<Current Measurement Accuracy>

DC	0.00 mA	-0.05 mA ~ 0.05 mA	(0.00 mA)
	20.00 mA	19.85 mA ~ 20.15 mA	(20.00 mA)

総合判定<Judgment> (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
----------------------------	--------------------------	-------------------------